CERTIFICATE OF ANALYSIS

Customer Name:	N/A	**Cylinder Number:**	GN0035333
Stock/Analyzer Tag #:	N/A	**Product Class:**	Primary Standard
Customer Reference:	N/A	**Cylinder Contents:**	220 cf @ 2000 psia
MESA Reference:	N/A	**Cylinder CGA:**	A047-HP-350/BR
Date of Certification:	N/A	**Analysis Method:**	GC-TCD
Recommended Shelf Life:	2 Years	**Preparation Method:**	Gravimetric

Component	Requested Concentration (2)	Reported Concentration (2,3)	Analytical Uncertainty
Methyl Ethyl Ketone	20 ppm	18.5 ppm	± 0.2 ppm
2,2-Dimethylbutane (Neohexane)	20 ppm	18.7 ppm	± 0.2 ppm
2,3-Dimethylbutane	20 ppm	18.9 ppm	± 0.2 ppm
3-Methyl-1-Pentene	20 ppm	18.8 ppm	± 0.2 ppm
2-Methylpentane (Isohexane)	20 ppm	18.6 ppm	± 0.2 ppm
2-Methyl-1-Butene	20 ppm	21.8 ppm	± 0.2 ppm
2-Methyl-2-Butene	20 ppm	18.6 ppm	± 0.2 ppm
2-Ethyl-1-Butene	20 ppm	19.1 ppm	± 0.2 ppm
Methyl T-Butyl Ether	20 ppm	19.0 ppm	± 0.2 ppm
Cyclohexane	20 ppm	18.5 ppm	± 0.2 ppm
Cyclopentane	20 ppm	19.0 ppm	± 0.2 ppm
Nitrogen	Balance	Balance	

Authorized Signature: _____

(1) The fill pressure shown on the COA is as originally quoted. The fill pressure measured by the customer may differ from the fill pressure originally quoted due to temperature effects, compressibility of the individual components when blended together in the cylinder, gauge accuracy or reduction in content volume before shipping as a result of samples withdrawn for laboratory QC necessary to ensure product quality.

(2) Unless otherwise stated, concentrations are given in molar units.

(3) Vapor pressure mixes are blended at a sufficiently low pressure so as to eliminate phase separation under most low temperature conditions encountered during transport or storage. However, it is generally recommended that cylinders containing vapor pressure restricted mixes be placed on the floor in a horizontal position and rolled back and forth to improve homogeneity of the gas phase mixture before being put into service.

Analytical Gas Standards are prepared and analyzed using combinations of NIST traceable weights, SRM's provided by NIST, or internal gas standards that have been verified for accuracy using procedures published by the US-EPA. Pure gases are analyzed and certified for purity using minor component Analytical Gas Standards prepared according to the methods specified above. Balances are calibrated to NIST test weights covered by NIST test number 822/278982-10, S/N 33071. Reference Certification #'s: 1000, 1004, 1005, & 1006.

MESA Specialty Gases & Equipment

division of MESA International Technologies, Inc.
2427 S. Anne St. • Santa Ana, California 92704 • USA
TEL: 714-434-7102 • FAX: 714-434-8006 • E-mail: mail@mesagas.com
On-line Catalog at www.mesagas.com